

**AMENDMENT NO. 1 DECEMBER 2016**  
**TO**  
**IS 8312 : 1977 SPECIFICATION FOR FORCEPS,**  
**INTRODUCING, ENDOTRACHEAL TUBE,**  
**MAGILL'S PATTERN**

*(Page 1, clause 3)* — Substitute the following for the existing clause:

**“3 Material**

Stainless steel bars conforming to Designation X20Cr13 or X30Cr13 of IS 6603 : 2001 ‘Stainless steel bars and flats (*first revision*)’.”